

TOPLEVEL PARAMETERS

Method Information For: D:\MSDCHEM\1\METHODS\MAN27.M

Method Sections To Run:

- (X) Save Copy of Method With Data
- () Pre-Run Cmd/Macro =
- (X) Data Acquisition
- (X) Data Analysis
- () Post-Run Cmd/Macro =

Method Comments:

Quantification du rapport Testosterone /Epitestosterone MSD20 injection en split

END OF TOPLEVEL PARAMETERS

INSTRUMENT CONTROL PARAMETERS

Sample Inlet: GC
 Injection Source: GC ALS
 Mass Spectrometer: Enabled

=====
 == 6890 GC METHOD ==
 =====

OVEN

Initial temp: 160 'C (On) Maximum temp: 325 'C
 Initial time: 0.00 min Equilibration time: 0.50 min

Ramps:

#	Rate	Final temp	Final time
1	4.00	255	0.00
2	30.00	300	2.75
3	0.0(Off)		

Post temp: 0 'C
 Post time: 0.00 min
 Run time: 28.00 min

FRONT INLET (UNKNOWN)

BACK INLET ()

Mode: Split
 Initial temp: 280 'C (On)
 Pressure: 175.0 kPa (On)
 Split ratio: 12.1:1
 Split flow: 11.7 mL/min
 Total flow: 15.2 mL/min
 Gas saver: Off
 Gas type: Helium

COLUMN 1

COLUMN 2

Method: MAN27.M

Sat Jul 22 13:31:13 2006

Page:1

USADA 0201

188